


Readjusted Xp low limit

Sep 20, 2017

motorx.adl

15:08:54



NEW  $\mu$ -stepping

Sep 22, 2017

motorx\_setup.adl

15:59:09

w/ Phytron @ 6

1 sin wave  
20

sipinkZp

(8idi:m80)

	USER	DIAL
High	18543.2500	[20000.0000]
Low	-21456.750	[-20000.000]
Encoder		Normal Backlash
Max. Speed (Rev/s)		1.000000
Speed (Rev/s)		1.000000 [1.000000]
Base speed (Rev/s)		0.250000
Accel time (s)		0.500000 [0.500000]
Backlash dist. ( microns )		50.000000
Move fraction		1.000000
Home Speed ( microns /s )		0.020000
Calibration		
Cal	Use	Set
Off	-1456.750	
		Variable
Dir	Pos	Neg
Resolution Readback		
Units:	microns	
Motor:	4000	Steps/Rev.
	1000.0000	microns/Rev.
Encoder:	0.250000	microns/Step
Readback:	0.000000	microns/Unit
RBV inLink:		
Use:	No	Yes
Use:	No	Yes
Retry		
Deadband	0.250000	
Max. retries	10	
Misc.		
Code version 6.10		
VME card#	5	
Precision	6	
NTM	NO	YES
NTM Factor	2	
More		

9/22 — Set top blade to microstepping  
went to top limit and set position  
to Table 2 position per Oxford. 165711mm

Sep 22, 2017

motorx.adl

15:18:08

Phytren set to 6

4000μstep/step **sipinkZp**

$\frac{1}{20}$

(8digit#80)

microns

16571.00000

16571.00000

TESTED ONE  
NEW cable  
by RAY

Labelled "C"


Repeated

to 16556.5

16463 16493, 16424  
w/ other  
cable

"D"

1000.000



Calib: Use Set

More

STOP

Scan

Ld

Go

Abort

Sep 22, 2017

motorx.adl

15:24:38

sipinkZp

(8idizm80)



microns

-3426.750000

-3426.750000

w/ step  
on 9/22/17  
  
Repeats w/  
cable  
"C" and  
"D"

300.0000



Calib: Use | Set

More

STOP

Scan Ld Go Abort

Sep 23, 2017

motorx.adl

23:24:16

# sipinkZm

(8idim79)

microns

-16025.00000

F16025.00001

1000.000



## Calib: Use Set

More

STOP

Scan

Ld

Go

Abort

Now  
μ-stepping

Zm 1  
20

# 6.

Used

Table 2

value

at low limit

i.e -16.025

mm

Adjusted

Zm top  
hard limit

HARD stop  
was too close..

On 9/25/17 Set  $X_p$  to zero as well  
as  $Z_m$

Sep 23, 2017

motorx.adl

23:33:20

